


86610530

Mangeshikar Uterine-Manipulator, Tenaculum



## Technical specifications

### Cleaning and Reprocessing

Reusable	Yes
Manual reprocessing	Yes, related to IFU
Washing machine	95 °C
Autoclave sterilization	max. 137 °C

### Approvals

Directive 93/42 EEC	Class IIa
Regulation (EU) 2017/745	N/A
FDA Approval	N/A

### Additional Product Information

Packaging unit	1
GTIN 14	04250418725477