



86610225

Mangeshikar Uterine-Manipulator, Shaft



## Technical specifications

### Cleaning and Reprocessing

|                         |                     |
|-------------------------|---------------------|
| Reusable                | Yes                 |
| Manual reprocessing     | Yes, related to IFU |
| Washing machine         | 95 °C               |
| Autoclave sterilization | max. 137 °C         |

### Approvals

|                          |           |
|--------------------------|-----------|
| Directive 93/42 EEC      | Class IIa |
| Regulation (EU) 2017/745 | N/A       |
| FDA Approval             | N/A       |

### Additional Product Information

|                |                |
|----------------|----------------|
| Packaging unit | 1              |
| GTIN 14        | 04250418725415 |